

Valeant Pharmaceuticals International, Inc.
400 Somerset Corporate Boulevard, Bridgewater, NJ 08807
Tel: 908-927-1400, Fax: 908-927-1856
www.valeant.com

**URGENT VOLUNTARY PRODUCT RECALL
AND FIELD CORRECTION TO LASEREDGE 1.5mm X 1.7mm Trapezoid Angled,
E7599 LOT MAKK060**

March 16, 2016

**Re: Notice of Dull Blades for LASEREDGE 1.5mm X 1.7mm Trapezoid Angled 6/Box
SKU#E7599 Lot MAKK060**

Product Details: 1.5mm X 1.7mm Trapezoid Angled Knife with handle (SKU# E7599 -6 per box)
manufactured by Angiotech (Surgical Specialties)


Product Name: LASEREDGE 1.5mm X 1.7mm Trapezoid Angled 6/Box SKU#E7599

Lot numbers of affected product shipped to you: MAKK060

Please Note: This recall is limited to E7599 Lot# MAKK060 and does not affect other lots
shipped to your facility

Dear Valued Customer,

This is to inform you of a medical device voluntary product recall and field correction involving
LASEREDGE 1.5mm X 1.7mm TRAPEZOID ANGLED 6/Box, SKU#E7599 Lot MAKK060. See below
the enclosed product label for ease in identifying the product at retail level.



Valeant Pharmaceuticals International wants to ensure that our products provide customer satisfaction and ease of use. It has come to our attention that some boxes of LASEREDGE 1.5mm X 1.7mm Trapezoid Angled Knives 6/Box, SKU#E7599 Lot MAKK060 have not been as sharp as previous lots of this product. Please review carefully the notes outlined in this letter

regarding your LASEREDGE 1.5mm X 1.7mm Trapezoid Angled Knives 6/Box, SKU#E7599 Lot MAKK060.

This action represents a voluntary product recall and we have notified the appropriate authority of this voluntary recall.

Please take the following steps to ensure the recalled product is handled appropriately:

1. **Quarantine the product:** According to our records, your facility may have a supply of LASEREDGE 1.5mm X 1.7mm Trapezoid Angled Knives 6/Box, SKU#E7599 Lot MAKK060. Please quarantine any unused boxes (full and partial) of the above product that you have in your facility. All unused boxes of LASEREDGE 1.5mm X 1.7mm Trapezoid Angled Knives 6/Box, SKU#E7599 Lot MAKK060 shall be returned to Valeant/B+L distribution center at the address below for a disposal.

Bausch+Lomb OPS B.V.
Koolhovenlaan 110
1119 NH Schiphol-Rijk
The Netherlands

2. **To obtain replacement product**

To obtain replacement product free of charge please acknowledge receipt of this notification by completing and returning the attached Recall and Field Correction Acknowledgement Form, with the date of this letter's receipt noted.

When completing this form be certain to do the following:

- Count the number of product packages in quarantine and note on the Form.
- Indicate your choice for receiving replacement product
- Fax the Form to 020 8781 0001 or email the form to uksurgorders@bausch.com

We sincerely apologize for any inconvenience and assure you that we are working diligently to resolve this issue in a timely manner. Please feel free to contact customer services with any questions or concerns: 020 8781 0000 or uksurgorders@bausch.com

PLEASE CUSTOMIZE FOR YOUR COUNTRY HERE AND ON FAX FORM.

Sincerely,

Eubelio Weston
RA DIRECTOR
MD, PS & CS VALEANT EMEA

Recall and Field Correction Acknowledgement Form

Product Details LASEREDGE 1.5mm X 1.7mm Trapezoid Angled Knives 6/Box, SKU#E7599 Lot MAKK060

This is to acknowledge receipt of the above referenced recall and field correction notification dated XXXXXXXX

Please confirm inventory levels of the affected cannulas with 4-digit lot numbers at your facility:

Product	Lot #	# Shipped	# used	# in inventory	Responsible person initials
E7599	MAKK060				

If you have inventory of affected E7599, Lot MAKK060 at your facility and would like to receive replacement product, please check the box below :

- I hereby certify that I have returned the existing inventory of E7599, Lot MAKK060 in our facility and ask that replacement product equal to the amount of quarantined product shown above, referencing PO# _____, be shipped to our facility.

Date

Name (Print)

Bausch + Lomb Account Number

Signature

Facility Name

Telephone Number

Please complete, sign and return this Form to:

Fax: XXXXXX

Email: XXXXXX

Confidentiality Agreement: The information contained in this facsimile message is privileged and confidential information intended for the use of the addressee listed above. If you are neither the intended recipient nor the employee or agent responsible for delivering this message to the intended recipient, you are hereby notified that any disclosure, copying, distribution or the taking of any action in reliance on the contents of the information is strictly prohibited. If you have received this in error, please immediately notify us by telephone to arrange for the return of the original document to us.