

**Aesculap AG  
Quality Management**

Postfach 40  
78501 Tuttlingen  
Germany

Contact: Christian Strobel  
  
Fon: +49 7461 95-31926  
Fax: +49 7461 95-1555  
Email: [vigilance\\_aag.de@aesculap.de](mailto:vigilance_aag.de@aesculap.de)  
Internet: <http://www.bbraun.com>  
  
Date: February 01, 2018

**Safety Notice – Product Recall**

**OM040R – DOYEN-COLLIN MOUTH GAG 120MM**

---

The Aesculap AG received knowledge about the possibility that the plastic sleeves of a DOYEN-COLLIN MOUTH GAG 120MM – OM040R could stick together after reprocessing.

The following image shows an affected product (see Figure 1)



Figure 1: OM040R – MOUTH GAG

It has been determined that the used plastic sleeves of the instrument do not meet the valid specification. Instead of the specified material silicone, the material PVC was used for the manufacturing of the sleeves. This could render affected instruments unusable after reprocessing.

Internal investigation conducted at the manufacturing plant revealed that the reported failure can be limited to the production period from February 2016 to October 2017.

**Chairman of Supervisory Board:**  
Prof. Dr. h.c. Ludwig Georg Braun

**Executive Board:**  
Dr. Joachim Schulz  
(Chairman)  
Dr. Jens von Lackum

**Corporate Office: Tuttlingen**  
Register Court: Stuttgart HRB 726261  
VAT reg. no. DE812160059  
  
WEEE-Reg.-No. DE 65109852

**Bank Account:**  
**Deutsche Bank AG Tuttlingen**  
BLZ 653 700 75 Konto 21 22 000 00  
IBAN DE44 6537 0075 0212 2000 00  
SWIFT / BIC DEUTDE33  
**Baden-Württembergische Bank**  
BLZ 600 501 01 Konto 487 1905  
IBAN DE31 6005 0101 0004 8719 05  
SWIFT / BIC SOLADEST

**Address:**  
Aesculap AG  
Am Aesculap-Platz  
78532 Tuttlingen  
Germany

An affected product OM040R can be clearly identified by two options:

Identification on the basis of the labelled production date **(from 02 16 to 10 17 inclusively)**.

Identification on the basis of the labelled encoding **(from 000330 to 000449 inclusively)**.

As soon as at least one of the two identification options is possible, it shall be assumed that the product is affected.



Figure 2: OM040R – LABELLING OF MOUTH GAG

The results of investigation revealed that there is no increased risk expected for patients who have been treated with the affected product.

According to our internal distribution information your facility received applicable units manufactured within the above mentioned production period. We kindly ask you to check if an affected instrument is currently in use at your facility.

**In case you have located an affected product:**

Please ensure that the involved products are no longer used.

Should you have an involved product, please return it with the attached "**Product Recall Form**" to

**Aesculap AG**  
**LRP**  
**Siegfried Schwarz**  
**Am Aesculap-Platz**  
**D-78532 Tuttlingen**

For any product-related requests, kindly do not hesitate to contact our product manager:

**Andreas Lauer**

**☎ + 49 7461 95 2479**

**☎ + 49 171 73 24 907**

**[andreas.lauer@aesculap.de](mailto:andreas.lauer@aesculap.de)**

**In case you could not locate any affected product:**

In the case you **do not have** any of the involved products, please send us the attached "**Feedback Form**" and tick as appropriate.

Please ensure that all users of the affected product are informed about this safety information in your organization. If you have distributed the products to a third party, please forward a copy of this information or inform the above mentioned contact person. The Competent Authority BfArM – Bundesinstitut für Arzneimittel und Medizinprodukte, has received a copy of this safety information.

We apologize for any inconvenience this may cause and thank you very much for your support.

With best regards,

Aesculap AG

i. V.

**Thorsten Barthelmes**  
Director Product Risk & Vigilance Management  
Safety Officer Medical Device

i. A.

**Kerstin Rothweiler**  
Team Leader Quality Management Vigilance  
Dpt. Safety Officer Medical Device

**FEEDBACK FORM / FSCA**

**OM040R - DOYEN-COLLIN MOUTH GAG 120MM**

Please send back this feedback form via fax or e-mail to:

**Department QMV**

**Fax +49 7461-95 1555**

**vigilance\_aag.de@aesculap.de**

☐ We do not have affected product(s).

☐ We will return affected product(s).

HOSPITAL \_\_\_\_\_ LOCATION \_\_\_\_\_

NAME \_\_\_\_\_ DEPARTMENT \_\_\_\_\_ PHONE \_\_\_\_\_

SIGNATURE \_\_\_\_\_ DATE \_\_\_\_\_



## PRODUCT RECALL

Hygienic condition:

☐ new good

☐ used decontaminated

☐ used not decontaminated

pos. no.	part no. article no.	serial / lot-no.	quantity	remark		
		n.a.				
		n.a.				
		n.a.				
		n.a.				
		n.a.				

### RETURN ADDRESS :

Aesculap AG  
LRP  
Siegfried Schwarz  
Am Aesculap-Platz  
D-78532 Tuttlingen – Germany

### ADRESS / SENDER:

### DATE / SIGNATURE :

Chairman of Supervisory Board:  
Prof. Dr. h.c. Ludwig Georg Braun

Executive Board:  
Dr. Joachim Schulz  
(Chairman)  
Dr. Jens von Lackum

Corporate Office: Tuttlingen  
Register Court: Stuttgart HRB 726261  
VAT reg. no. DE812160059  
  
WEEE-Reg.-No. DE 65109852

Bank Account:  
**Deutsche Bank AG Tuttlingen**  
BLZ 653 700 75 Konto 21 22 000 00  
IBAN DE44 6537 0075 0212 2000 00  
SWIFT / BIC DEUTDE33  
**Baden-Württembergische Bank**  
BLZ 600 501 01 Konto 487 1905  
IBAN DE31 6005 0101 0004 8719 05  
SWIFT / BIC SOLADEST

Address:  
Aesculap AG  
Am Aesculap-Platz  
78532 Tuttlingen  
Germany