



Waldemar Link GmbH & Co. KG
Implantate und
Chirurgische Instrumente

Barkhausenweg 10
22339 Hamburg · Germany
Telefon: +49 40 53995-0
Fax: +49 40 5386929
E-Mail: info@linkhh.de
Internet: linkorthopaedics.com

Waldemar Link GmbH & Co. KG, Postfach 63 05 52, 22315 Hamburg

IK 330 200 672

Datum

01st of March 2018

IMPORTANT SAFETY INFORMATION FOR USERS

Ansprechpartner/in

Notification of a product recall

Affected product:	Alignment guide 55° of TOP II Impactor Handle
Item-Number:	Subunit 15°/55° of the article 132-150/03
Lot-Numbers:	All [Excluded are alignment guides 55° with an own engraved article no.: 183-150/05]
LINK® Reference-Number:	R-2018-01

Dear Sir or Madame,

With this letter we inform you about a recall of the above mentioned product. It is noticed that the alignment guide 55° of the impactor handle is not conform with the specification. As you were supplied with the affected products, please read carefully the information on the following pages.

We would be grateful if you could return the fax reply to us in any case until the **19th of March 2018**, as documentation of the recall, even if you have none of the listed products in stock or if these products do not exhibit the defect in question.

In conclusion, we apologize for any inconveniences caused and thank you for your understanding in this matter.

Yours sincerely,
WALDEMAR LINK GmbH & Co. KG

Kai Lemke
Safety Officer