

Field Safety Notice PR 1457388

23 May 2018

Field Safety Notice

FSCA identifier: Product Field Corrective Action – **PR 1457388**

Type of Action: Field Safety Corrective Action

Description: Update of Instructions for Use (IFU) for non-sterile Instruments and trials

Affected Product(s): Surgical Instruments and Trials

Lot Numbers: All Surgical Instruments and Trials

Dear Customer,

On 23rd May 2018, Stanmore Implants Worldwide Ltd. has initiated a voluntary Field Safety Corrective Action related to the IFUs of non-sterile, reusable Surgical Instruments and Trials. This letter serves to notify users to a change in the Surgical Instruments and Trials IFUs, list potential hazards associated with the steam sterilization of the devices and list risk mitigation factors.

Issue

It was identified that the IFUs supplied with non-sterile instruments and trials stated UK steam sterilisation parameters (134-137°C for 3 minutes - pre-vacuum), while the validated sterilisation parameters were US parameters (132°C for 4 minutes - porous load).

To remediate the non-validated sterilisation parameters gap and to align the sterilisation parameters with standard UK requirements, Stanmore Implants Worldwide Ltd. performed a steam sterilisation validation of all reusable instrument kits and trials to standard UK sterilisation parameters (minimum 3 minutes at 134-137°C – pre-vacuum) to achieve a sterility assurance level (SAL) of not less than 10⁻¹² at full cycle. The validation also determined that steam penetration during sterilisation is improved where the polymeric cap of the General Impactor (Part No. imgenimp) and its variant, Long General Impactor (Part No. imlgimp) are partially un-threaded (approximately 2mm) ensuring that the device still fits in its specified location within the tray.

IFU documentation has been updated accordingly (See Appendix I for updated IFUs for Surgical Instruments and Trials).

The changes made in these IFUs are as follows:

1. Sterilisation parameters in Section 3 have been changed:

Sterilisation parameters outlined in previous IFUs

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Porous load	132°C	4 minutes (minimum)	30 minutes (minimum)

Sterilisation parameters in the updated IFUs

<i>Method</i>	<i>Cycle Type</i>	<i>Temperature</i>	<i>Exposure Time</i>	<i>Drying Time</i>
<i>Steam</i>	<i>Pre-vacuum</i>	<i>134-137°C</i>	<i>3 minutes (minimum)</i>	<i>30 minutes (minimum)</i>

2. Additional text has been added to Section 3:
 - a. *Partially unthread the polymeric cap of the General Impactor and its variant (approximately 2mm) while still ensuring that it fits in its specified location within the tray, to improve steam penetration during sterilisation.*
 - b. *Ensure the polymeric cap is re-tightened on the General Impactor and its variant prior to use.*
 - c. *Ensure all components are dry prior to use.*

Potential Hazards/Harms

The potential hazard(s) of not unthreading the General Impactor and its variant may include:

- Infectious agents remain on the device due to cleaning, disinfection and inadequate sterilization

Risk Mitigation:

1. Disassembly of the polymeric cap of the general impactor and its variant during cleaning and disinfection.
2. Partially unthreading the polymeric cap of the general impactor and its variant during sterilisation.

Actions Needed

1. Please inform users of this FSCA and forward this notice to all those individuals who need to be aware within your organization.
2. Hospitals: Complete and sign the enclosed acknowledgment Form and fax or email a copy to the Customer Services at +44 (0) 20 8953 0617 or Mets.requests@stammoreimplants.com

Our records indicate that you have received the above referenced device. It is our responsibility to ensure that customers who may have received this affected device also receive this important communication.

Please assist us in meeting our regulatory obligation by faxing back the attached Business Reply Form within 5 days.

Stanmore Implants Worldwide Limited maintains its commitment to developing, manufacturing and marketing the highest quality products for surgeons and patients. We apologize for any inconvenience this Field Safety Corrective Action may create and appreciate your cooperation with our request.

If you have any further enquiries, please contact your local Sales Representative.

Yours Sincerely,



Dervillia Murphy
Director, Quality Assurance and Regulatory Compliance
210 Centennial Avenue,
Elstree, Hertfordshire, WD6 3SJ,
United Kingdom
+44 20 8238 6500
Dervillia.Murphy@stryker.com
Business hours: 9am – 5pm (GMT)

Attachments:

1. Acknowledgement Form
2. Appendix I - Updated Surgical Instruments and Trials IFUs
 - TF04-6 Issue 11: Surgical Instruments and Trials Instructions for Use
 - DM01-8 Issue 10: Custom Made Prosthetic Replacement Instructions for Use
 - QL041 Issue 04: JTS Extendible Distal Femoral Implant Instructions for Use
 - QL042 Issue 05: JTS Extendible Proximal Tibial Implant Instructions for Use

**STANMORE IMPLANTS WORLDWIDE LIMITED
FIELD SAFETY CORRECTIVE ACTION ACKNOWLEDGMENT FORM**

23 May 2018

NAME:

ADDRESS:

CITY, STATE ZIP,

POSTCODE:

FSCA identifier: Product Field Corrective Action **PR 1457388**

Type of Action: Field Safety Corrective Action

Description: Update of Instructions for Use (IFU) for non-sterile Instruments and Trials

Affected Product(s): Surgical Instruments and Trials

Product Information

Lot: All surgical Instruments and Trials

I confirm that I have received the IFUs appended as Appendix I and have reviewed the UK sterilisation parameters in Section 3 along with the additional instructions regarding the sterilisation of General Impactor and its variant.

Customer
(Signature)

Date

Customer Name
(PRINT)

Please email this signed and dated form to Mets.requests@stammoreimplants.com

Appendix I - Updated Surgical Instruments and Trials IFUs

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

JTS[®] Non-Invasive Extendible Proximal Tibia Implant

IFU - Instructions for Use

Please Read in Conjunction with the Surgical Technique and the Operation Drawing Before Commencing Surgery

1. Implant description

The Stanmore Implants Worldwide JTS[®] Extendible Proximal Tibia Implant is manufactured in accordance with an approved prescription for a named patient. It **MUST NOT** be used for any other patient.

The patient name is detailed in the operation drawing supplied with the implant.

JTS[®] Extendible implant intended use/indications for use

The JTS[®] Extendible implant is indicated for cemented or cementless limb sparing procedures where radical resection and replacement of the proximal tibia is required with the following conditions:

Patients suffering from severe arthropathy of the knee that does not respond to any conservative therapy or better alternative surgical treatment;

Surgical intervention for severe trauma, revision arthroplasties, failed previous prostheses and/or oncology indications; and malignant diseases (e.g. osteogenic sarcoma).

The JTS[®] Extendible Proximal Tibia Implant and its components are for single use only.

2. Warnings and precautions:



Stanmore Implants Worldwide JTS[®] Extendible Proximal Tibial Implant is manufactured in accordance with an approved prescription for a named patient. It must not be used for any other patient.



Stanmore Implants Worldwide JTS[®] Extendible Proximal Tibial Implant must not be used with products from other manufacturers. Different manufacturers have different tolerances and therefore a mismatch could lead to failure of the implant.



The JTS[®] Extendible Proximal Tibial Implant is supplied sterilised and must not be steam sterilised. Contact Stanmore Implants for advice.



If cortical bone screws are to be used to fix the extra-cortical plate in place, these must be of a suitable size to fit through the plate (nominally $\varnothing 4.5\text{mm}$) and made from implant grade titanium.



The JTS[®] Extendible Proximal Tibial Implant is MRI UNSAFE and neither the implant nor a patient with an implant in place must be taken into an MRI environment.



The Stanmore Implants Worldwide JTS[®] Extendible Proximal Tibial Implant is for Single Use Only.



If the packaging of parts marked sterile has been compromised or is damaged do not use.



An operation instruction supplied with the implant contains patient specific information to aid the implantation



Do not steam sterilise the JTS[®] Extendible Proximal Tibial Implant as this will damage the internal magnet and it will not extend.



The surgical technique instruction must be read prior to carrying out any surgical procedure.



Once implant has been fully inserted and secured with axle and circlip or axle and axle-cap, remove retaining clip located on the femoral component. Failure to do so will result in malfunction of the device and may lead to future surgical intervention.

























The patient should be cautioned to limit activities and protect the replaced joint from unreasonable stresses and follow the instructions of the physician with respect to follow-up care and treatment. The patient should be closely monitored if a change at the operative site has been detected. The possibility of deterioration of the joint should be evaluated and possible revision surgery considered.



All polyethylene components are supplied sterile and must not be re-sterilised if the sterility of the component has been compromised an alternative sterile component must be used

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

- | | | | |
|---|--|---|---|
|  | <p>The patient should be warned of surgical risks and made aware of possible adverse effects. The patient should be warned that the prosthesis does not replace normal healthy bone. That the prosthesis can break or become damaged as a result of certain activity or trauma has a finite expected service life and may need to be replaced at some time in the future. The patient should also be advised of other risks that the surgeon believes should be disclosed.</p> |  | <p>For any items that are supplied sterile check to ensure that they are still within their shelf life prior to use.</p> |
|  | <p>Hitting or dropping the JTS® Extendible Proximal Tibial Implant can result in demagnetising the magnets in the growing section which will result in being unable to grow the implant.</p> |  | <p>If there is more than six months between supply of the JTS® Extendible Proximal Tibial Implant and the implantation, further scans must be produced and reviewed against the design by Stanmore Implants Worldwide as this could result in a possible mis-match due to changes in bone geometry.</p> |
|  | <p>Particular care and attention should be taken with respect to all bearing surfaces, tapers and ceramic coated surfaces.</p> |  | <p>If revision of the implant is required, this should be carried out by a suitably qualified person. There are no special techniques required for the revision of a JTS® Extendible Proximal Tibial Implant. The revision procedure would use the same surgical procedure and tooling required for implantation</p> |
|  | <p>If the HA coating is contaminated or damaged do not try to clean. Remove from use and contact Stanmore Implants Worldwide for advice.</p> |  | <p>Before any action is taken to grow the JTS® Extendible Proximal Tibial Implant, the JTS Drive Unit operations manual must be read and understood</p> |
|  | <p>The HA coated components are not to be cemented in place.</p> |  | <p>Specialised instruments are required to assemble and disassemble the components. Using other instruments may damage the components. Instruments can become damaged or may even fracture after repeated use or if used with excessive force. Instruments should be inspected prior to surgery and not used if damaged or worn excessively</p> |
|  | <p>Improper use or mishandling of the components can result in damage to one or more of the components reducing the in-service life of the implant.</p> |  | <p>Specialised instruments are required to assemble and disassemble the components. Using other instruments may damage the components. Instruments can become damaged or may even fracture after repeated use or if used with excessive force. Instruments should be inspected prior to surgery and not used if damaged or worn excessively</p> |
|  | <p>The surgical instruments are supplied not sterile. Sterilise before use in accordance with the instructions provided.</p> |  | <p>Before any action is taken to grow the JTS® Extendible Proximal Tibial Implant, the operation Drawing must be consulted as this gives details of the direction of magnetic field rotation for each individual implant.</p> |
|  | <p>Surgical instruments must never be placed in physiological saline solution as contact leads to pitting and corrosion.</p> |  | <p>Before any action is taken to grow the JTS® Extendible Proximal Tibial Implant, the operation Drawing must be consulted as this gives details of the direction of magnetic field rotation for each individual implant.</p> |
|  | <p>Improper cleaning tools (e.g. overlarge brushes, metal brushes) can cause damage. DO NOT use scouring agents or steel wool.</p> |  | <p>CT scans obtained with a metal implant inside the patient do produce artefacts and therefore tissue definition in the adjacent area is compromised and this must be taken into account when reviewing any results.</p> |
|  | <p>Ensure detergent solutions operate within a pH range of 6.0-8.0.</p> |  | <p>CT scans obtained with a metal implant inside the patient do produce artefacts and therefore tissue definition in the adjacent area is compromised and this must be taken into account when reviewing any results.</p> |
|  | <p>Any deviation from recommended sterilisation methods must be validated by the user.</p> | | |
|  | <p>The recommended sterilisation method is only valid with sterilisation equipment that is maintained, calibrated and operated in accordance with local regulatory guidelines.</p> | | |
|  | <p>Ensure steam quality meets acceptable standards to prevent damage and discolouration.</p> | | |

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

Indications and Complications

2.1. Indications:

- Primary bone tumours
- Secondary tumours arising in bone
- Non-neoplastic conditions affecting the shafts of long bone
- Failed joint replacements
- Failed massive replacements

The JTS® Extendible Proximal Tibial implant is indicated for cemented or cementless procedures where radical resection and replacement of the proximal tibia is required.

2.2. Contra-indications:

Absolute contra-indications include

- Infection and sepsis

Relative contra-indications include

- Long delay between manufacturing and insertion of a patient specific implant may result in significant mismatch due to possible changes in bone geometry.
- Inadequate or incomplete soft tissue coverage.
- Uncooperative or unwilling patient or patient unable to follow instructions
- Foreign body sensitivity. Where materials sensitivity occurs seek advice with respect to testing
- Obesity
- Vascular disorder, neuromuscular disorders or muscular dystrophy

2.3. Patient Selection:

Factors that should be considered are:

- Resection of neoplastic or diseased bone
- At risk from pathological fracture
- Pain relief and improved function
- Ability of patient to willingly follow instructions and undergo rehabilitation

2.4. Possible adverse effects:

There is a range of potential adverse reactions; these may include:

- Patient sensitivity to implant materials which may ultimately require removal of the device
- Infection which may require temporary or permanent removal of the device
- Discolouration of the adjacent tissues may occur
- Fretting between metal parts is possible under certain circumstances

2.5. Intraoperative and early postoperative complications:

These may include:

- Temporary or permanent nerve damage
- Damage to blood vessels
- Haematoma
- Cardiovascular disorders
- Pulmonary embolism
- Myocardial infarction or venous thrombosis
- Delayed wound healing
- Infection
- Loosening
- Varus and valgus deformity
- Dislocation

2.6. Late postoperative complications:

These may include:

- Loosening
- Bone resorption
- Bone fracture
- Fatigue fracture of metal components
- Wear of components due to misalignment or excessive loading
- Inadequate range of movement
- Infection
- Metal sensitivity or allergic reaction

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

3. Cleaning and Sterilisation

3.1. Implant sterilisation:

IMPORTANT



THE JTS® EXTENDIBLE PROXIMAL TIBIAL IMPLANT IS SUPPLIED STERILE **DO NOT STEAM STERILISE THE IMPLANT AS THIS WILL AFFECT ITS FUNCTION**



The implant has been sterilised using 25 to 40kGy gamma irradiation.



If the packaging of parts marked sterile has been compromised or is damaged **DO NOT** use.



If the sterility of the component has been compromised, an alternative sterile component **MUST** be used.

For all items that are not marked sterile (this includes all surgical instruments) the procedure in Section 3.2 must be followed.

3.2. Orthopaedic Instruments and Loaner sets:

All instruments are supplied NON-STERILE and must be cleaned and sterilised before use and before being returned to Stanmore Implants in accordance with instructions in this section.

a. **Information on methods**

Listed here are validated methods for disinfection, cleaning and sterilisation for user reprocessing of Stanmore Implants Worldwide Limited, orthopaedic Instruments and loaner sets. As far as possible, it is recommend that decontamination is performed using an automatic washing-disinfector utilising thermal disinfection, followed by steam sterilisation.

b. **Preparation for cleaning**

At the point of use, remove all excess visible soil and keep the reusable medical devices moist until cleaning. Any device capable of disassembly must be disassembled before cleaning. Ensure that thorough cleaning and rinsing is performed as soon as possible. Manual cleaning may be undertaken either by immersion or non-immersion techniques dependent upon the device construction but these instructions relate solely to a validated automated cleaning cycle.

c. **Cleaning Equipment and Agents**

Water

Water quality is an important consideration in all cleaning steps. Demineralised or deionised water is recommended as this can help prevent discolouration and staining.

Detergents and Cleaning Agents

Only suitable detergents and cleaning agents within a range of pH 6.0-8.0 may be used.

Ultrasonic Cleaner

Ultrasonic cleaning is suitable for an especially thorough, mild cleaning of heavily soiled, and other difficult to access locations such as joints, crevices and channels.

Cleaning Equipment

Non-abrasive low linting cloth and general purpose cleaning brushes.

d. **Product Cleaning Guidelines – Automatic**

The following validated cleaning instruction is based on general product features which present challenges to the cleaning process and not on specific products. The instructions presented are based on an automated method using a washer-disinfector and have been validated.

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

In the case of non-metals, unless otherwise instructed, reprocess following these guidelines.

Equipment and reagents:

Washer/disinfector (Medisafe SI Auto Workstation), Enzyme based detergent (3E-Zyme).

Procedure:

- i. Load the instruments into washer/disinfector ensuring that the basket is not overloaded and that hinges are open and cannulation holes are able to drain.
- ii. Run cycle (key stages):
 - Cold pre-wash: 3 minutes, < 35°C
 - Main wash: 13 minutes, 43°C
 - Pre Rinse: 2 minutes, 20°C
 - Disinfection fill
 - Thermal disinfection: 2 minutes, 80-85°C
 - Air purge dry: 2 minutes
- iii. When the cycle is complete remove the contents from the washer for inspection.
- iv. Inspect each instrument for cleanliness. If soiling is still evident repeat the cleaning cycle.

e. Steam Sterilisation

IMPORTANT



THE JTS® EXTENDIBLE IMPLANT IS SUPPLIED STERILE **DO NOT STEAM STERILISE THE IMPLANT AS THIS WILL AFFECT ITS FUNCTION**



Partially unthread the polymeric cap of the General Impactor (approximately 2mm) while still ensuring that it fits in its specified location within the tray, to improve steam penetration during sterilisation.

The following sterilisation process should be used for those items supplied non-sterile.

It is recommended that these items are sterilised using prevacuum or porous load high temperature steam sterilisation

These devices must be placed in suitably wrapped porous membrane packaging for the sterilisation process (i.e. central supply wrap, autoclave bags, paper/polyethylene pouches etc.).

The suggested cycles, based on recommended industry standards and regulatory guidance consist of the following:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Prevacuum	134-137°C	3 minutes (minimum)	30 minutes (minimum)

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Porous load	132 +/- 2°C (266-273°F)	4 minutes (minimum)	30 minutes (minimum)

Stanmore Implants has validated the above recommended sterilisation cycles. Other sterilisation methods and cycles may also be suitable. Individuals or hospitals are advised, to validate whichever method is considered appropriate for their organisation. Ethylene Oxide (EtO) and cold sterilisation techniques are not recommended.

Ensure that the polymeric cap is re-tightened on the General Impactor prior to use.

Ensure all components are dry prior to use.

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE

3.3. Re-Sterilisation



If the JTS® Extendible Proximal Tibial Implant or polyethylene components require re-sterilisation they must be returned to Stanmore Implants. They **MUST NOT** be steam sterilised.



Orthopaedic Instruments and loaner sets should be cleaned and re-sterilised in accordance with the instructions in Section 3.2

JTS EXTENDIBLE PROXIMAL TIBIA IMPLANT INSTRUCTIONS FOR USE



	<p>Company Information</p> <p>Manufacturer Stanmore Implants Worldwide Ltd 210 Centennial Avenue Centennial Park Elstree Hertfordshire WD6 3SJ United Kingdom</p> <p>Tel: +44 (0) 20 8238 6500</p> <p>Fax: +44 (0) 20 8953 0617</p>	<p>Contact Information</p> <p>If further information is required on Stanmore Implants' devices or instrumentation please contact the Design Services Office:</p> <p>Tel: +44 (0) 20 8238 6500 Fax: +44 (0) 20 8953 0617</p> <p>E-mail: designgroup@stryker.com</p>
---	--	--

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE

JTS[®] Non-Invasive Extendible Distal Femoral Implant

IFU - Instructions for Use

Please Read in Conjunction with the Surgical Technique and the Operation Drawing Before Commencing Surgery

1. Implant description

The Stanmore Implants Worldwide JTS[®] Extendible Distal Femoral implant is manufactured in accordance with an approved prescription for a named patient. It **MUST NOT** be used for any other patient.

The patient name is detailed in the operation drawing supplied with the implant.

JTS[®] Extendible implant intended use/indications for use

The JTS[®] Extendible implant is indicated for cemented and cementless limb sparing procedures where radical resection and replacement of the distal femur is required with the following conditions:

Patients suffering from severe arthropathy of the knee that does not respond to any conservative therapy or better alternative surgical treatment;

Surgical intervention for severe trauma, revision arthroplasties, failed previous prostheses and/or oncology indications; and malignant diseases (e.g. osteogenic sarcoma).

The JTS[®] Extendible Distal Femoral Implant and its components are for single use only.

2. Warnings and precautions:



Stanmore Implants Worldwide JTS[®] Extendible Distal Femoral Implant is manufactured in accordance with an approved prescription for a named patient. It must not be used for any other patient.



Stanmore Implants Worldwide JTS[®] Extendible Distal Femoral Implant must not be used with products from other manufacturers. Different manufacturers have different tolerances and therefore a mismatch could lead to failure of the implant.



The JTS[®] Extendible Distal Femoral Implant is supplied sterilised and must not be steam sterilised. Contact Stanmore Implants for advice.



If cortical bone screws are to be used to fix the extra-cortical plate in place, these must be of a suitable size to fit through the plate (nominally $\varnothing 4.5\text{mm}$) and made from implant grade titanium.



The JTS[®] Extendible Distal Femoral Implant is MRI UNSAFE and neither the implant nor a patient with an implant in place must be taken into an MRI environment.



The Stanmore Implants Worldwide JTS[®] Extendible Distal Femoral Implant is for Single Use Only.



If the packaging of parts marked sterile has been compromised or is damaged do not use.



An operation instruction supplied with the implant contains patient specific information to aid the implantation



Do not steam sterilise the JTS[®] Extendible Distal Femoral Implant as this will damage the internal magnet and it will not extend.



The operation technique instruction must be read prior to carrying out any surgical procedure.



Once implant has been fully inserted and secured with axle and circlip or axle and axle-cap, remove retaining clip located on the femoral component. Failure to do so will result in malfunction of the device and may lead to future surgical intervention.



The patient should be cautioned to limit activities and protect the replaced joint from unreasonable stresses and follow the instructions of the physician with respect to follow-up care and treatment. The patient should be closely monitored if a change at the operative site has been detected. The possibility of deterioration of the joint should be evaluated and possible revision surgery considered.



All polyethylene components are supplied sterile and must not be re-sterilised if the sterility of the component has been compromised an alternative sterile component must be used

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE



The patient should be warned of surgical risks and made aware of possible adverse effects. The patient should be warned that the prosthesis does not replace normal healthy bone. That the prosthesis can break or become damaged as a result of certain activity or trauma has a finite expected service life and may need to be replaced at some time in the future. The patient should also be advised of other risks that the surgeon believes should be disclosed.



For any items that are supplied sterile check to ensure that they are still within their shelf life prior to use.



If there is more than six months between supply of the JTS[®] Extendible Distal Femoral Implant and the implantation, further scans must be produced and reviewed against the design by Stanmore Implants Worldwide as this could result in a possible mis-match due to changes in bone geometry.



Hitting or dropping the JTS[®] Extendible Distal Femoral Implant can result in demagnetising the magnets in the growing section which will result in being unable to grow the implant.



If revision of the implant is required, this should be carried out by a suitably qualified person. There are no special techniques required for the revision of a JTS[®] Extendible Distal Femoral Implant. The revision procedure would use the same surgical procedure and tooling required for implantation



Particular care and attention should be taken with respect to all bearing surfaces, tapers and ceramic coated surfaces.



Before any action is taken to grow the JTS[®] Extendible Distal Femoral Implant, the JTS Drive Unit operations manual must be read and understood



If the HA coating is contaminated or damaged do not try to clean. Remove from use and contact Stanmore Implants Worldwide for advice.



The HA coated components are not to be cemented in place.



Specialised instruments are required to assemble and disassemble the components. Using other instruments may damage the components. Instruments can become damaged or may even fracture after repeated use or if used with excessive force. Instruments should be inspected prior to surgery and not used if damaged or worn excessively



Improper use or mishandling of the components can result in damage to one or more of the components reducing the in-service life of the implant.



The Surgical instruments are supplied not sterile. Sterilise before use in accordance with the instructions provided.



Before any action is taken to grow the JTS[®] Extendible Distal Femoral Implant, the operation Drawing must be consulted as this gives details of the direction of magnetic field rotation for each individual implant.



Surgical instruments must never be placed in physiological saline solution as contact leads to pitting and corrosion.



Improper cleaning tools (e.g. overlarge brushes, metal brushes) can cause damage. **DO NOT** use scouring agents or steel wool.



CT scans obtained with a metal implant inside the patient do produce artefacts and therefore tissue definition in the adjacent area is compromised and this must be taken into account when reviewing any results.



Ensure detergent solutions operate within a pH range of 6.0-8.0.



Any deviation from recommended sterilisation methods must be validated by the user.



The recommended sterilisation method is only valid with sterilisation equipment that is maintained, calibrated and operated in accordance with local regulatory guidelines.



Ensure steam quality meets acceptable standards to prevent damage and discolouration.

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE

Indications and Complications

2.1. Indications:

- Primary bone tumours
- Secondary tumours arising in bone
- Non-neoplastic conditions affecting the shafts of long bone
- Failed joint replacements
- Failed massive replacements

The JTS® Extendible Distal Femoral implant is indicated for cemented and cementless procedures where radical resection and replacement of the distal femur is required.

2.2. Contra-indications:

Absolute contra-indications include

- Infection and sepsis

Relative contra-indications include

- Long delay between manufacturing and insertion of a patient specific implant may result in significant mismatch due to possible changes in bone geometry.
- Inadequate or incomplete soft tissue coverage.
- Uncooperative or unwilling patient or patient unable to follow instructions
- Foreign body sensitivity. Where materials sensitivity occurs seek advice with respect to testing
- Obesity
- Vascular disorder, neuromuscular disorders or muscular dystrophy

2.3. Patient Selection:

Factors that should be considered are:

- Resection of neoplastic or diseased bone
- At risk from pathological fracture
- Pain relief and improved function
- Ability of patient to willingly follow instructions and undergo rehabilitation

2.4. Possible adverse effects:

There is a range of potential adverse reactions; these may include:

- Patient sensitivity to implant materials which may ultimately require removal of the device
- Infection which may require temporary or permanent removal of the device
- Discolouration of the adjacent tissues may occur
- Fretting between metal parts is possible under certain circumstances

2.5. Intraoperative and early postoperative complications:

These may include:

- Temporary or permanent nerve damage
- Damage to blood vessels
- Haematoma
- Cardiovascular disorders
- Pulmonary embolism
- Myocardial infarction or venous thrombosis
- Delayed wound healing
- Infection
- Loosening
- Varus and valgus deformity
- Dislocation

2.6. Late postoperative complications:

These may include:

- Loosening
- Bone resorption
- Bone fracture
- Fatigue fracture of metal components
- Wear of components due to misalignment or excessive loading
- Inadequate range of movement
- Infection
- Metal sensitivity or allergic reaction

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE

3. Cleaning and Sterilisation

3.1. Implant sterilisation:

IMPORTANT



THE JTS® EXTENDIBLE DISTAL FEMORAL IMPLANT IS SUPPLIED STERILE **DO NOT STEAM STERILISE THE IMPLANT AS THIS WILL AFFECT ITS FUNCTION**



The implant has been sterilised using 25 to 40kGy gamma irradiation.



If the packaging of parts marked sterile has been compromised or is damaged **DO NOT** use.



If the sterility of the component has been compromised, an alternative sterile component **MUST** be used.

For all items that are not marked sterile (this includes all surgical instruments) the procedure in Section 3.2 must be followed.

3.2. Orthopaedic Instruments and Loaner Sets:

All instruments are supplied NON-STERILE and must be cleaned and sterilised before use and before being returned to Stanmore Implants in accordance with instructions in this section.

a. **Information on methods**

Listed here are validated methods for disinfection, cleaning and sterilisation for user reprocessing of Stanmore Implants Worldwide Limited, orthopaedic Instruments and loaner sets. As far as possible, it is recommend that decontamination is performed using an automatic washing-disinfector utilising thermal disinfection, followed by steam sterilisation.

b. **Preparation for cleaning**

At the point of use, remove all excess visible soil and keep the reusable medical devices moist until cleaning. Any device capable of disassembly must be disassembled before cleaning. Ensure that thorough cleaning and rinsing is performed as soon as possible. Manual cleaning may be undertaken either by immersion or non-immersion techniques dependent upon the device construction but these instructions relate solely to a validated automated cleaning cycle.

c. **Cleaning Equipment and Agents**

Water

Water quality is an important consideration in all cleaning steps. Demineralised or deionised water is recommended as this can help prevent discolouration and staining.

Detergents and Cleaning Agents

Only suitable detergents and cleaning agents within a range of pH 6.0-8.0 may be used.

Ultrasonic Cleaner

Ultrasonic cleaning is suitable for an especially thorough, mild cleaning of heavily soiled, and other difficult to access locations such as joints, crevices and channels.

Cleaning Equipment

Non-abrasive low linting cloth and general purpose cleaning brushes.

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE

d. Product Cleaning Guidelines – Automatic

The following validated cleaning instruction is based on general product features which present challenges to the cleaning process and not on specific products. The instructions presented are based on an automated method using a washer-disinfector and have been validated.

In the case of non-metals, unless otherwise instructed, reprocess following these guidelines.

Equipment and reagents:

Washer/disinfector (Medisafe SI Auto Workstation), Enzyme based detergent (3E-Zyme).

Procedure:

- i. Load the instruments into washer/disinfector ensuring that the basket is not overloaded and that hinges are open and cannulation holes are able to drain.
- ii. Run cycle (key stages):
 - Cold pre-wash: 3 minutes, < 35°C
 - Main wash: 13 minutes, 43°C
 - Pre Rinse: 2 minutes, 20°C
 - Disinfection fill
 - Thermal disinfection: 2 minutes, 80-85°C
 - Air purge dry: 2 minutes
- iii. When the cycle is complete remove the contents from the washer for inspection.
- iv. Inspect each instrument for cleanliness. If soiling is still evident repeat the cleaning cycle.

e. Steam Sterilisation

IMPORTANT



THE JTS® EXTENDIBLE IMPLANT IS SUPPLIED STERILE **DO NOT STEAM STERILISE THE IMPLANT AS THIS WILL AFFECT ITS FUNCTION**



Partially unthread the polymeric cap of the General Impactor (approximately 2mm) while still ensuring that it fits in its specified location within the tray, to improve steam penetration during sterilisation.

The following sterilisation process should be used for those items supplied non-sterile.

It is recommended that these items are sterilised using prevacuum or porous load high temperature steam sterilisation

These devices must be placed in suitably wrapped porous membrane packaging for the sterilisation process (i.e. central supply wrap, autoclave bags, paper/polyethylene pouches etc.).

The suggested cycles, based on recommended industry standards and regulatory guidance consist of the following:

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Prevacuum	134-137°C	3 minutes (minimum)	30 minutes (minimum)

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Porous load	132 +/- 2°C (266-273°F)	4 minutes (minimum)	30 minutes (minimum)

Ensure the polymeric cap is re-tightened on the General Impactor prior to use.

Ensure all components are dry prior to use.

Stanmore Implants has validated the above recommended sterilisation cycles. Other sterilisation methods and cycles may also be suitable. Individuals or hospitals are advised, to validate whichever method is considered appropriate for

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE

their organisation. Ethylene Oxide (EtO) and cold sterilisation techniques are not recommended.

3.3. **Re-Sterilisation**



If the JTS[®] Extendible Distal Femoral Implant or polyethylene components require re-sterilisation they must be returned to Stanmore Implants. They **MUST NOT** be steam sterilised.



Orthopaedic Instruments and loaner sets should be cleaned and re-sterilised in accordance with the instructions in Section 3.2

JTS EXTENDIBLE DISTAL FEMORAL IMPLANT INSTRUCTIONS FOR USE



	<p>Company Information</p> <p>Manufacturer Stanmore Implants Worldwide Ltd 210 Centennial Avenue Centennial Park Elstree Hertfordshire WD6 3SJ United Kingdom</p> <p>Tel: +44 (0) 20 8238 6500 Fax: +44 (0) 20 8953 0617</p>	<p>Contact Information</p> <p>If further information is required on Stanmore Implants' devices or instrumentation please contact the Design Services Office:</p> <p>Tel: +44 (0) 20 8238 6500 Fax: +44 (0) 20 8953 0617</p> <p>E-mail: designgroup@stryker.com</p>
---	--	--

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**

Content	Section	Pages
Warnings and Precautions	1	1
Indications and contraindications	2	3
Patient selection	2	3
Possible adverse effects	2	3
Intraoperative and early postoperative complications	2	3
Late postoperative complications	2	4
Instrument sterilisation	3	4
Instrument Cleaning	4	5
Re-sterilisation	5	6
Declaration of contamination status	Appendix 1	7

Implant description:

SIW “Custom made” implants are manufactured in accordance with an approved prescription for a named patient. It **MUST NOT** be used for any other patient.

Section 1



WARNINGS and PRECAUTIONS



SIW Custom made implants are manufactured in accordance with an approved prescription for a named patient. It must not be used for any other patient.



If the packaging of parts marked sterile has been compromised or is damaged do not use.



All plastic components are supplied sterile and must not be re-sterilised if the sterility of the component has been compromised an alternative sterile component must be used.



Stanmore Implants Worldwide custom implants must not be used with products from other manufacturers unless it is specifically authorised. Different manufacturers have different tolerances and therefore a mismatch could lead to failure of the implant. An operation instruction supplied with the implant contains any such authorisation.



The Stanmore Implants Worldwide custom implant are for Single Use Only.

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.



An operation instruction supplied with the implant contains patient specific information to aid the implantation.



The operation technique instruction must be read prior to carrying out any surgical procedure.



The patient should be cautioned to limit activities and protect the replaced joint from unreasonable stresses and follow the instructions of the physician with respect to follow-up care and treatment. The patient should be closely monitored if a change at the operative site has been detected. The possibility of deterioration of the joint should be evaluated and possible revision surgery considered.



The patient should be warned of surgical risks and made aware of possible adverse effects. The patient should be warned that the prosthesis does not replace normal healthy bone, that the prosthesis can break or become damaged as a result of certain activity or trauma, has a finite expected service life and may need to be replaced at some time in the future. The patient should also be advised of other risks that the surgeon believes should be disclosed.



Improper use or mishandling of the components can result in damage to one or more of the components reducing the in-service life of the implant.



Particular care and attention should be taken with respect to all bearing surfaces, tapers and ceramic coated surfaces.



If the HA coating is contaminated or damaged do not try to clean. Remove from use and contact Stanmore Implants Worldwide for advice.



The Surgical instruments are supplied not sterile. Clean and sterilise before use in accordance with the instructions provided



Surgical instruments must never be placed in physiological saline solution as contact leads to pitting and corrosion.



Improper cleaning tools (e.g. overlarge brushes, metal brushes) can cause damage. **DO NOT** use scouring agents or steel wool.



Ensure detergent solutions operate within a pH range of 6.0-8.0.



Any deviation from recommended sterilisation methods must be validated by the user.



The recommended sterilisation method is only valid with sterilisation equipment that is maintained, calibrated and operated in accordance with local regulatory guidelines.



Ensure steam quality meets acceptable standards to prevent damage and discolouration



For any items that are supplied sterile check to ensure that they are still within their shelf life prior to use.

If there is more than two months between supply custom implant and the implantation,

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**



further scans must be produced and reviewed against the design by Stanmore Implants Worldwide as this could result in a possible mis-match due to changes in bone geometry.



Specialised instruments are required to assemble and disassemble the components. Using other instruments may damage the components. Instruments can become damaged or may even fracture after repeated use or if used with excessive force. Instruments should be inspected prior to surgery and not used if damaged or worn excessively

Section 2

Indications:

- Primary bone tumours
- Secondary tumours arising in bone
- Non-neoplastic conditions affecting the shafts of long bone
- Failed joint replacements
- Failed massive replacements

Contraindications:

- Absolute contra-indications include
 - Infection and sepsis
- Relative contra-indications include
 - Long delay between manufacturing and insertion of a custom-made implant may result in significant mismatch due to possible changes in bone geometry
 - Inadequate or incomplete soft tissue coverage
 - Uncooperative or unwilling patient or patient unable to follow instruction
 - Foreign body sensitivity. Where materials sensitivity occurs seek advice with respect to testing
 - Obesity
 - Vascular disorder, neuromuscular disorders or muscular dystrophy

Patient Selection:

Factors that should be considered are

- Resection of neoplastic or diseased bone
- At risk from pathological fracture
- Pain relief and improved function
- Ability of patient to willingly follow instructions and under go rehabilitation

Possible adverse effects:

There are a range of potential adverse reactions these may include

- Patient sensitivity to implant materials which may ultimately require removal of the device
- Infection which may require temporary or permanent removal of the device
- Discolouration of the adjacent tissues may occur
- Fretting between metal parts is also possible under certain circumstances

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**

Intraoperative and early postoperative complications

These may include

- Temporary or permanent nerve damage
- Damage to blood vessels
- Haematoma
- Cardiovascular disorders
- Pulmonary embolism
- Myocardial infarction or venous thrombosis
- Delayed wound healing
- Infection
- Loosening
- Varus and valgus deformity
- Dislocation

Late postoperative complications

These may include

- Loosening
- Bone resorption
- Bone fracture
- Fatigue fracture of metal components
- Wear of components due to misalignment or excessive loading
- Inadequate range of movement
- Infection
- Metal sensitivity or allergic reaction

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**

Cleaning and Sterilisation

All Instruments must be cleaned and sterilised prior to returning to Stanmore Implants Worldwide in accordance with the procedures defined in sections 3 and 4 and the form in appendix 1 must be completed and returned with any items being sent back to Stanmore Implants.

Section 3

INSTRUMENT STERILISATION

IMPORTANT



All instruments **MUST** be cleaned prior to sterilisation (see section 4).



Partially unthread the polymeric cap of the General Impactor (approximately 2mm) while still ensuring that it fits in its specified location within the tray, to improve steam penetration during sterilization.



If the implant is marked as STERILE on the packaging. It has been sterilised using 25 to 40KGy gamma irradiation **DO NOT** re-sterilise.



All components that are supplied sterile will be marked on the packaging as STERILE. They have been sterilised using 25 to 40KGy gamma irradiation **DO NOT** re-sterilise.



If the packaging of parts marked sterile has been compromised or is damaged **DO NOT** use.



All plastic components are supplied sterile and **MUST NOT** be re-sterilised. If the sterility of the component has been compromised, an alternative sterile component must be

For all items that are not marked sterile the following procedure should be followed:

The following sterilisation process should be used for instruments supplied non-sterile

The instruments are recommended to be sterilised using prevacuum or porous load, high temperature steam sterilisation

These devices must be placed in suitably wrapped porous membrane packaging for the sterilisation process (i.e. central supply wrap, autoclave bags, paper/plastic pouches etc.).

The suggested cycles, based on recommended industry standards and regulatory guidance consist of the following:

Method:	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Prevacuum	134-137m	3 minutes (minimum)	30 minutes (minimum)

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**

Method:	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Porous Load	134 +/-2°C	4 minutes (minimum)	30 minutes (minimum)

Stanmore Implants has validated the above recommended sterilisation cycles. Other sterilisation methods and cycles may also be suitable. Individuals or hospitals are advised, to validate whichever method is considered appropriate for their organisation. Ethylene Oxide (EtO) and cold sterilisation techniques are not recommended.

Ensure the polymeric cap is re-tightened on the General Impactor prior to use.

Ensure all components are dry prior to use.

Section 4

ORTHOPAEDIC INSTRUMENTS AND LOANER SETS

All instruments are supplied cleaned but are NON-STERILE and must be cleaned and sterilised before use and before being returned to Stanmore Implants in accordance with instructions in section 3 and 4 and the form in appendix 1 must be completed.

a. Information on methods

Listed here are validated methods for disinfection, cleaning and sterilisation for user reprocessing of Stanmore Implants Worldwide Limited, Orthopaedic Instruments and Loaner sets. As far as possible, it is recommended that decontamination is performed using an automatic washing-disinfector utilising thermal disinfection, followed by steam sterilisation.

b. Preparation for cleaning

At the point of use, remove all excess visible soil and keep the reusable medical devices moist until cleaning. Any device capable of disassembly must be disassembled before cleaning. Ensure that thorough cleaning and rinsing is performed as soon as possible. Manual cleaning may be undertaken either by immersion or non-immersion techniques dependent upon the device construction but these instructions relate solely to a validated automated cleaning cycle.

c. Cleaning Equipment and Agents

Water	Water quality is an important consideration in all cleaning steps. Demineralised or deionised water is recommended as this can help prevent discolouration and staining.
Detergents and Cleaning Agents	Only suitable detergents and cleaning agents within a range of pH 6.0-8.0 may be used.
Ultrasonic Cleaner	Ultrasonic cleaning is suitable for an especially thorough, mild cleaning of heavily soiled, and other difficult to access locations such as joints, crevices and channels.
Cleaning Equipment	Non-abrasive low linting clothe and general purpose cleaning brushes.

CUSTOM MADE PROSTHETIC REPLACEMENT INSTRUCTIONS FOR USE

**PLEASE READ THIS IN CONJUNCTION WITH THE SUPPLIED OPERATION
TECHNIQUE INSTRUCTION AND THE OPERATION DRAWING.**

d. Product Cleaning Guidelines – Automatic

The following validated cleaning instruction is based on general product features which present challenges to the cleaning process and not on specific products. The instructions presented are based on an automated method using a washer-disinfector and have been validated. In the case of non-metals, unless otherwise instructed, reprocess following these guidelines.

Equipment and reagents:

Washer/disinfector (Medisafe SI Auto Workstation), Enzyme based detergent (3E-Zyme).

Procedure:

- i. Load the instruments into washer/disinfector ensuring that the basket is not overloaded and that hinges are open and cannulation holes are able to drain.
- ii. Run cycle (key stages):
 - Cold pre-wash: 3 minutes, < 35°C
 - Main wash: 13 minutes, 43°C
 - Pre Rinse: 2 minutes, 20°C
 - Disinfection fill
 - Thermal disinfection: 2 minutes, 80-85°C
 - Air purge dry: 2 minutes
- iii. When the cycle is complete remove the contents from the washer for inspection.
- iv. Inspect each instrument for cleanliness. If soiling is still evident repeat the cleaning cycle.

Section 5


RE-STERILISATION

Plastic parts **MUST NOT** be re-sterilised.

If an implant requires re-sterilisation please contact SIW for advice.

Re-sterilise using an autoclave as per section 3. Re-sterilisation and validation of the autoclave is the responsibility of the hospital and not SIW or its agent.

Orthopaedic Instruments and Loaner sets should be cleaned and re-sterilised in accordance with the instructions in section 3 and 4

	<p>Company Information</p> <p>Manufacturer</p> <p>Stanmore Implants Worldwide Ltd 210 Centennial Avenue Centennial Park Elstree Hertfordshire WD6 3SJ United Kingdom Tel +44 (0) 208 2386500 Fax +44 (0) 208 9537443</p>	<p>Contact Information</p> <p>If further information is required on Stanmore Implants devices or instrumentation please contact the Design Office:</p> <p>Tel: (+44) 020 89541402 Fax: (+44) 020 89530617</p> <p><u>E-mail: designgroup@stryker.com</u></p>

Declaration of Contamination Status

This form is only to be used for the return of Instruments used with a custom implants for standard product please use appendix in form TF04-06.

Local forms can be supplied as an alternative as long as all the required information is supplied.

Product Description				Product Identification			
From				Address			
Contact Name							
Emergency contact number				Contact e-mail			
Have any of the items been contaminated		Yes *	No	Don't Know	Please circle		
* State type of contamination: Blood, Body fluids, or any other hazard							
Have the items been decontaminated		Yes †	No ‡	Don't Know	Please circle		
† Was the process in accordance with the information given DM01/8-7 July 09 sections 3 and 4					Yes	No	Please circle
IF NO please provide details of what cleaning and sterilisation process was used							
‡ Please explain why the items have not been decontaminated							
Signature of person completing the form				Date			
Print Name		Job Title					
By signing this form you are confirming that all of the information is correct and accurate to the best of your knowledge at the time of approval.							
<p>CONTAMINATED ITEMS MUST NOT BE RETURNED WITHOUT THE PRIOR AGREEMENT AND KNOWLEDGE OF STANMORE IMPLANTS WORLDWIDE.</p> <p>THEY MUST BE RETURNED SUITABLY PACKAGED AND IDENTIFIED.</p>							












SURGICAL INSTRUMENTS AND TRIALS INSTRUCTIONS FOR USE

PLEASE READ THIS IN CONJUNCTION WITH SURGICAL PROCEDURE

<u>Content</u>	<u>Section</u>	<u>Pages</u>
Warnings and Precautions	1	1
Trials and Instrument Cleaning	2	2
Trials and instrument sterilisation	3	3
Re-sterilisation	4	4
Declaration of contamination status	Appendix 1	5
Release Note	Appendix 2	6

Section 1

WARNINGS and PRECAUTIONS

-  **DO NOT** implant the trial implants
-  Stanmore Implants Worldwide trial implants must not be used with products from other manufacturers.
-  The surgical procedure must be read prior to carrying out any surgical procedure.
-  For instructions in the use of the trial instruments refer to the surgical planning guide provided.
-  Improper use or mishandling of the components can result in damage to one or more of the components, or improper selection of implants.
-  The surgical instruments and trials are supplied **NON-STERILE**. Clean and sterilise before use, in accordance with the instructions provided (see Section 2 and 3.)
-  Surgical instruments must never be placed in physiological saline solution as contact leads to pitting and corrosion.
-  Improper cleaning tools (e.g. overlarge brushes, metal brushes) can cause damage. **DO NOT** use scouring agents or steel wool.
-  Ensure detergent solutions operate within a pH range of 6.0-8.0.
-  Any deviation from recommended sterilisation methods must be validated by the user.
-  The recommended sterilisation method is only valid with sterilisation equipment that is maintained, calibrated and operated in accordance with local regulatory guidelines.

SURGICAL INSTRUMENTS AND TRIALS INSTRUCTIONS FOR USE

PLEASE READ THIS IN CONJUNCTION WITH SURGICAL PROCEDURE



Ensure steam quality meets acceptable standards to prevent damage and discolouration



Do not sterilise using either Ethylene Oxide (EtO) or cold sterilisation techniques.

Cleaning and Sterilisation

All Instruments and trials must be cleaned and sterilised prior to returning to Stanmore Implants Worldwide in accordance with the procedures defined in Sections 2 and 3 and the form in Appendix 1 must be completed and returned with any items being sent back to Stanmore Implants.

Section 2

TRIALS and INSTRUMENTS CLEANING INSTRUCTIONS

All instruments are supplied cleaned but are NON-STERILE and must be cleaned and sterilised before use and before being returned to Stanmore Implants in accordance with the instructions in Sections 2 and 3 and the form in Appendix 1 must be completed.

a. Information on methods

Listed here are validated methods for disinfection, cleaning and sterilisation for user reprocessing of Stanmore Implants Worldwide Limited, Orthopaedic Instruments and Trials sets. As far as possible, it is recommend that decontamination is performed using an automatic washing-disinfector utilising thermal disinfection, followed by steam sterilisation.

b. Preparation for cleaning

At the point of use, remove all excess visible soil and keep the reusable medical devices moist until cleaning. Any device capable of disassembly must be disassembled before cleaning. Ensure that thorough cleaning and rinsing is performed as soon as possible. Manual cleaning may be undertaken either by immersion or non-immersion techniques dependent upon the device construction but these instructions relate solely to a validated automated cleaning cycle.

c. Cleaning Equipment and Agents

Water	Water quality is an important consideration in all cleaning steps. Demineralised or deionised water is recommended as this can help prevent discolouration and staining.
Detergents and Cleaning Agents	Only suitable detergents and cleaning agents within a range of pH 6.0-8.0 may be used.
Ultrasonic Cleaner	Ultrasonic cleaning is suitable for an especially thorough, mild cleaning of heavily soiled, and other difficult to access locations such as joints, crevices and channels.
Cleaning Equipment	Non-abrasive low linting clothe and general purpose cleaning brushes.

SURGICAL INSTRUMENTS AND TRIALS INSTRUCTIONS FOR USE

PLEASE READ THIS IN CONJUNCTION WITH SURGICAL PROCEDURE

d. Product Cleaning Guidelines – Automatic

The following validated cleaning instruction is based on general product features which present challenges to the cleaning process and not on specific products. The instructions presented are based on an automated method using a washer-disinfector and have been validated.

In the case of non-metals, unless otherwise instructed, reprocess following these guidelines.

Equipment and reagents:

Washer/disinfector (Medisafe SI Auto Workstation), Enzyme based detergent (3E-Zyme).

Procedure:

- i. Load the instruments into washer/disinfector ensuring that the basket is not overloaded and that hinges are open and cannulation holes are able to drain.
- ii. Run cycle (key stages):
 - Cold pre-wash: 3 minutes , < 35°C
 - Main wash: 13 minutes, 43°C
 - Pre Rinse: 2 minutes, 20°C
 - Disinfection fill
 - Thermal disinfection: 2 minutes, 80-85°C
 - Air purge dry: 2 minutes
- iii. When the cycle is complete remove the contents from the washer for inspection.
- iv. Inspect each instrument for cleanliness. If soiling is still evident repeat the cleaning cycle.

Section 3

TRIALS and INSTRUMENT STERILISATION

IMPORTANT



All trials and instruments **MUST** be cleaned prior to sterilisation (see section 2)



Partially unthread the polymeric cap of the General Impactor (approximately 2mm) while still ensuring that it fits in its specified location within the tray, to improve steam penetration during sterilisation.

The following sterilisation process should be used for trials and instrumentation.

The trials and instrumentation are recommended to be sterilised using prevacuum or porous load, high temperature steam sterilisation (air removal via pulsed pre-vacuum method)

These devices must be placed in suitably wrapped porous membrane packaging for the sterilisation process (i.e. central supply wrap, autoclave bags, paper/plastic pouches etc.).

The suggested cycles, based on recommended industry standards and regulatory guidance consist of the following:

SURGICAL INSTRUMENTS AND TRIALS INSTRUCTIONS FOR USE

PLEASE READ THIS IN CONJUNCTION WITH SURGICAL PROCEDURE

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Prevacuum	134-137°C	3 minutes (minimum)	30 minutes (minimum)

Method	Cycle Type	Temperature	Exposure Time	Drying Time
Steam	Porous load	132°C	4 minutes (minimum)	30 minutes (minimum)

Stanmore Implants has validated the above recommended sterilisation cycles. Other sterilisation methods and cycles may also be suitable. Individuals or hospitals are advised, however to validate whichever method is considered appropriate for their organisation. Ethylene Oxide (EtO) and cold sterilisation techniques are not recommended.

Ensure the polymeric cap is re-tightened on the General Impactor prior to use.
Ensure all components are dry prior to use.

Section 4

RE- STERILISATION

Orthopaedic instruments and trials should be cleaned and re-sterilised in accordance with the instructions in section 2 and 3.

 <p>Company Information Manufacturer Stanmore Implants Worldwide Ltd 210 Centennial Avenue Centennial Park Elstree Hertfordshire WD6 3SJ United Kingdom Tel +44 (0) 208 2386500 Fax +44 (0) 208 9537443</p> <p>CE 0120</p>	<p>Contact Information</p> <p>If further information is required on Stanmore Implants devices or instrumentation please contact the Design Office:</p> <p>Tel: (+44) 020 8954 1402 Fax: (+44) 020 8953 0167 E-mail: designgroup@stryker.com</p>
--	--

Declaration of Contamination Status

This form is only to be used for the return of Trials and Instruments used with a METS modular system product. Local forms can be supplied as an alternative as long as all the required information is supplied.

Product Description		Product Identification		
From		Address		
Contact Name		Emergency contact number		
Contact e-mail		Have any of the items been contaminated		
Yes *	No	Don't Know	Please circle	
* State type of contamination: Blood, Body fluids, or any other hazard				
Have the items been decontaminated		Yes †	No ‡	Don't Know
		Please circle		
† Was the process in accordance with the information given TF04-06 Issue 8 sections 2 and 3		Yes	No	Please circle
IF NO please provide details of what cleaning and sterilisation process was used				
‡ Please explain why the items have not been decontaminated				
Signature of person completing the form		Date		
Print Name	Job Title			
By signing this form you are confirming that all of the information is correct and accurate to the best of your knowledge at the time of approval.				
ANY CONTAMINATED INSTRUMENTS OR TRIALS <u>MUST NOT</u> BE RETURNED WITHOUT THE PRIOR AGREEMENT AND KNOWLEDGE OF STANMORE IMPLANTS WORLDWIDE. THE INSTRUMENTS OR TRIALS MUST BE RETURNED SUITABLY PACKAGED AND IDENTIFIED.				

METS RELEASE NOTE

***METS* MODULAR SYSTEM**

CRC Number: _____

SURGICAL INSTRUMENTS AND TRIALS: (Decontamination Certificate)

Special Instruments/ Trials included with the above system

Yes No

These instruments/trials were previously used in a surgical invasive procedure and were exposed to blood, body fluids or pathological samples. These items were subsequently cleaned and sterilised prior to return to Stanmore Implants Worldwide By (Insert organisation name or leave blank)

In accordance with the validated process detailed in TF04-6 Issue 8.

After inspection all of the instruments and trails are cleaned and washed in accordance with the process defined in TF04-6 Issue 8.

I declare that I have taken all reasonable steps to ensure the accuracy of the above information

Signature

Date

Print Name

If further information is required on Stanmore Implants devices or instrumentation please contact the Design Office:

Tel: (+44) 020 8954 1402

Fax: (+44) 020 8953 0167

E-mail: designgroup@stryker.com