



■ **HUM GmbH**, Zum Pier 79, D-44536 Lünen

# Important Safety Notice

## Update of the instructions for use AEROpart<sup>®</sup> - O<sub>2</sub>-cannula B

**Sender:** HUM Gesellschaft für Homecare und Medizintechnik mbH

**Adressee:** Users, operators and trade partners

### Affected medical devices

AEROpart<sup>®</sup> - O<sub>2</sub>-cannula B, oxygen cannula B, adults, cannula with ear pieces and 1,5 m (5 ft.) supply tubing

| Article no. | Article description   |
|-------------|---|
| HSB01-B     | AEROpart <sup>®</sup> - O <sub>2</sub> -cannula B, oxygen cannula B, adults, cannula with ear pieces and 1,5 m (5 ft.) supply tubing, straight nasal tips |

■ **Deutschland | Germany**  
HUM Gesellschaft für Homecare und Medizintechnik mbH  
Zum Pier 79 | D-44536 Lünen  
FON +49 (0)231 880885-0  
FAX +49 (0)231 880885-58  
E-Mail: sales@hum-online.de  
Internet: www.hum-online.de

■ **Büro China | Office China**  
HUM Homecare and Medical Technology  
A45 F9 North Tower, Poly V Mansion  
Xiancun RD 11, Pearl River New City, Tianhe District  
Guangzhou, 510627 Guangdong, China  
FON +86 20 66691812  
FAX +86 20 66691812 999  
E-mail: china-office@hum-online.de

■ **IK-Nr. | Code of the institution**  
330 592 869  
Ust.-ID-Nr. | Sales Tax Identification Number  
DE180373438  
Registergericht | Commercial Register Number  
Dortmund HRB 17477  
Geschäftsführer | General Manager  
Dietmar Uhlenbruch

■ **Bankverbindung | Bank Account**  
Sparkasse Dortmund  
BLZ 440 501 99  
Konto 491 006 250  
IBAN: DE28 4405 0199 0491 0062 50  
BIC/SWIFT: DORTDE33XXX  
WEEE-Reg.-Nr: DE 55961085

■ **HUM GmbH**, Zum Pier 79, D-44536 Lünen

Dear Ladies and Gentlemen,

In order to ensure the safety of our medical devices, we request that you implement the corrective measures described below as a part of your obligation to cooperate in accordance with the medical device implementation legislation, so that users can continue to use our medical devices safely on patients.

### **Problem Description**

When checking our instructions for use for the product **AEROpart® - O<sub>2</sub>-cannula B**, it was found that important user information is missing. It is extremely important that users are made aware of the information that has been added to the new instruction for use.

- The device must be connected and activated by qualified personnel.
- The product is intended for single patient use only.
- The connection of the O<sub>2</sub>-cannulas is described.
- The checking of the connections is described.
- The „keep away from fire“ notice is added.
- A note on disposal has been added.
- The manufacturer's symbol has been included in the symbol explanation.

### **Risk for the Patient**

No patient injuries have been reported to us so far. Nevertheless, we decided to take this voluntary measure because the safety of your patients is our top priority.

■ **Deutschland | Germany**  
HUM Gesellschaft für Homecare  
und Medizintechnik mbH  
Zum Pier 79 | D-44536 Lünen  
FON +49 (0)231 880885-0  
FAX +49 (0)231 880885-58  
E-Mail: [sales@hum-online.de](mailto:sales@hum-online.de)  
Internet: [www.hum-online.de](http://www.hum-online.de)

■ **Büro China | Office China**  
HUM Homecare and Medical Technology  
A45 F9 North Tower, Poly V Mansion  
Xiancun RD 11, Pearl River New City, Tianhe District  
Guangzhou, 510627 Guangdong, China  
FON +86 20 66691812  
FAX +86 20 66691812 999  
E-mail: [china-office@hum-online.de](mailto:china-office@hum-online.de)

■ **IK-Nr. | Code of the institution**  
330 592 869  
Ust.-ID-Nr. | Sales Tax Identification Number  
DE180373438  
Registergericht | Commercial Register Number  
Dortmund HRB 17477  
Geschäftsführer | General Manager  
Dietmar Uhlenbruch

■ **Bankverbindung | Bank Account**  
Sparkasse Dortmund  
BLZ 440 501 99  
Konto 491 006 250  
IBAN: DE28 4405 0199 0491 0062 50  
BIC/SWIFT: DORTDE33XXX  
WEEE-Reg.-Nr: DE 55961085

■ **HUM GmbH**, Zum Pier 79, D-44536 Lünen

## Corrective Measures

The enclosed instructions for use has missing application and safety instructions for using the **AEROp<sup>®</sup>art - O<sub>2</sub>-cannula B.**

Therefore the following corrective measures must be carried out:

- Please check your inventory and block all affected products immediately until the measures are implemented.
- Completely fill out the attached response form, even if you no longer have the affected products in stock.
- If you have passed on affected items to a third party, please ensure that your customers are promptly informed of this recall by forwarding this letter. Your customers also have to fill out the reply form and send it back to us.

Based on the data on the answer sheet, we will immediately send you a corresponding number of updated instructions for use and polybags.

- Please pack the unopened product in the sent polybag. Insert the new instructions for use **ID 885974 Rev. 6.0 | 10.01.2022** into the new outer packaging.

Please implement these measures **IMMEDIATELY!**

We require a response from each notified customer. Therefore, please send us the completed reply form until February 28, 2022.

We feel sorry for the inconvenience caused to you by this action and ask for your understanding for this measure.

Please contact us at the contact provided if you have any questions regarding this letter and its implementation.

**Contact:** Melissa Korf, Tel.: +49 231 88 08 85-63, E-Mail: [korf@hum-online.de](mailto:korf@hum-online.de)

Best regards

HUM  
Gesellschaft für Homecare  
und Medizintechnik mbH



Dietmar Uhlenbruch  
CEO



i.A. Melissa Korf  
Quality Management

■ Deutschland | Germany  
HUM Gesellschaft für Homecare  
und Medizintechnik mbH  
Zum Pier 79 | D-44536 Lünen  
FON +49 (0)231 880885-0  
FAX +49 (0)231 880885-58  
E-Mail: [sales@hum-online.de](mailto:sales@hum-online.de)  
Internet: [www.hum-online.de](http://www.hum-online.de)

■ Büro China | Office China  
HUM Homecare and Medical Technology  
A45 F9 North Tower, Poly V Mansion  
Xiancun RD 11, Pearl River New City, Tianhe District  
Guangzhou, 510627 Guangdong, China  
FON +86 20 66691812  
FAX +86 20 66691812 999  
E-mail: [china-office@hum-online.de](mailto:china-office@hum-online.de)

■ IK-Nr. | Code of the institution  
330 592 869  
Ust.-ID-Nr. | Sales Tax Identification Number  
DE180373438  
Registergericht | Commercial Register Number  
Dortmund HRB 17477  
Geschäftsführer | General Manager  
Dietmar Uhlenbruch

■ Bankverbindung | Bank Account  
Sparkasse Dortmund  
BLZ 440 501 99  
Konto 491 006 250  
IBAN: DE28 4405 0199 0491 0062 50  
BIC/SWIFT: DORTDE33XXX  
WEEE-Reg.-Nr: DE 55961085

■ **HUM GmbH**, Zum Pier 79, D-44536 Lünen

## Response Form

*Exchange of the instructions for use* **AEROpart® - O<sub>2</sub>-cannula B**

Please fill out this reply form and send it to us by e-mail, fax or post to:

E-Mail: [qualitymanagement@hum-online.de](mailto:qualitymanagement@hum-online.de)

Fax: +49 231 880 885 58

Post: HUM Gesellschaft für Homecare und Medizintechnik mbH  
Zum Pier 79  
44536 Lünen

- We hereby confirm receipt of the letter. We have read and understood the content and will implement all measures immediately. All users and operators of the product and other persons to be informed have received the letter.**

**Please fill our completely in block letters:**

Sender: \_\_\_\_\_

Address: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Contact person: \_\_\_\_\_

Telephone No: \_\_\_\_\_ Fax No: \_\_\_\_\_

Date/Signature: \_\_\_\_\_

- We still have \_\_\_\_\_ pieces of the AEROpart® - O<sub>2</sub>-cannula B in stock.**
- We no longer have this product in stock.**

■ **Deutschland | Germany**  
HUM Gesellschaft für Homecare  
und Medizintechnik mbH  
Zum Pier 79 | D-44536 Lünen  
FON +49 (0)231 880885-0  
FAX +49 (0)231 880885-58  
E-Mail: [sales@hum-online.de](mailto:sales@hum-online.de)  
Internet: [www.hum-online.de](http://www.hum-online.de)

■ **Büro China | Office China**  
HUM Homecare and Medical Technology  
A45 F9 North Tower, Poly V Mansion  
Xiancun RD 11, Pearl River New City, Tianhe District  
Guangzhou, 510627 Guangdong, China  
FON +86 20 66691812  
FAX +86 20 66691812 999  
E-mail: [china-office@hum-online.de](mailto:china-office@hum-online.de)

■ **IK-Nr. | Code of the institution**  
330 592 869  
Ust.-ID-Nr. | Sales Tax Identification Number  
DE180373438  
Registergericht | Commercial Register Number  
Dortmund HRB 17477  
Geschäftsführer | General Manager  
Dietmar Uhlenbruch

■ **Bankverbindung | Bank Account**  
Sparkasse Dortmund  
BLZ 440 501 99  
Konto 491 006 250  
IBAN: DE28 4405 0199 0491 0062 50  
BIC/SWIFT: DORTDE33XXX  
WEEE-Reg.-Nr: DE 55961085