

## Urgent Field Safety Notice

Date: 10 October 2017

Reference: CC17-029

**Action:** Advisory

Our records indicate that your facility received one or more of the products subject to this Field Safety Notice. The table below gives a full list of the affected products.

Affected Product	Part Number	Lot number
Yelloport 10.5/12.5 Valve Single-Use	580004S	70126

### Description of the problem:

This Field Safety Notice gives notification of a labelling error on the product listed above.

The expiry date on some cartons (secondary packaging) has been identified with the incorrect date. The label states the expiry date as the same date as the date of manufacture, 2013-12.

The correct expiry date for Lot number 70126 is 2018-12.


The label applied to the device (primary packaging) displays the correct expiry date (see picture 1).

### Picture 1



The label applied to the carton (secondary packaging) was found to be labelled with the incorrect expiry date (see picture 2).

**Picture 2**



A thorough investigation has been carried out and therefore this issue does not impact the quality of the product or pose a risk to patient safety.

**The product may be used up to the expiry date of 2018-12 as per the label on the primary packaging**

**Customer Instructions:**

- Please read this notice
- Complete the attached acknowledgement, even if you no longer have any of the product.
- Please return the completed end user and distributor acknowledgement to [shazia.chaudhry@surginno.co.uk](mailto:shazia.chaudhry@surginno.co.uk) within 5 working days.
- Please pass this notice onto all users of the product listed within your organisation.
- Surgical Innovations is not requesting return or removal of these products as a result of this Field Safety Notice. This notice is advisory.

Although Surgical Innovations is not recalling any of the product listed above, your National Competent Authority will be notified of this Field Safety Notice.

If you have any questions please contact Surgical Innovations on the contact details below.

Contact reference person:

Shazia Chaudhry  
Quality Assurance and Regulatory Manager  
Surgical Innovations Ltd  
Clayton Wood House  
6 Clayton Wood Bank  
Leeds  
LS16 6QZ  
England  
Tel: + 44 (0) 113 2307597  
Fax: + 44 (0) 113 2307598  
Email: [shazia.chaudhry@surginno.co.uk](mailto:shazia.chaudhry@surginno.co.uk)

Thank you for your consideration.