



3. Please make sure that this information is observed in your institution until all necessary internal measures have been completed.
4. To verify receipt of this urgent safety information we need the included **answer form** which we kindly ask you to return to **fax number 07043 354467**, or via e-mail to **marek.rast@richard-wolf.com by October 01, 2018**. Please fill in this form even if the product is no longer in your stock. With the answer form you confirm receipt of this urgent safety information and avoid getting further reminder letters from Richard Wolf GmbH.
5. Inform Richard Wolf GmbH about any adverse events.

**Do not use the affected Bronchoscope tubes listed.**

Please return the products to Richard Wolf GmbH. We will perform the corresponding corrective measures on the affected tubes and send them back to you. Please send the products to:

Richard Wolf GmbH  
Customerservice  
Pforzheimer Straße 32  
75438 Knittlingen

The competent national authorities (amongst others the Federal Institute for Drugs and Medical Devices in Bonn) have been informed about this **urgent safety information**.

Your contact persons

for questions about the handling:

Herr Marek Rast  
Head of Technical Service  
Tel.: +49 7043 35 1046  
Fax: +49 7043 35 4467  
E-Mail: marek.rast@richard-wolf.com

for safety-relevant inquiries:

Herr Heiko Seider-Biedermann  
Deputy Safety Officer for Medical Products  
Tel.: +49 7043 35 4112  
Fax: +49 7043 35 4300  
E-Mail: heiko.seider-biedermann@richard-wolf.com

We apologize for the inconvenience caused by this measure and thank you on behalf of Richard Wolf GmbH for your support ensuring a timely implementation.

Kind regards,

**Richard Wolf GmbH**

  
p.p. Marek Rast  
Dept. Manager Technical Service

  
p.p. Heiko Seider-Biedermann  
Deputy Safety Officer for Medical Products

Appended: answer form, listing of affected products