



Waldemar Link GmbH & Co. KG
Implantate und
Chirurgische Instrumente

Barkhausenweg 10
22339 Hamburg · Germany
Telefon: +49 40 53995-0
Fax: +49 40 5386929
E-Mail: info@linkhh.de
Internet: linkorthopaedics.com

IK 330 200 672

Waldemar Link GmbH & Co. KG, Postfach 63 05 52, 22315 Hamburg

Datum

23 November 2018


Ansprechpartner/in

IMPORTANT SAFETY INFORMATION FOR USERS

Notification of a product recall

Affected product: MP Prosthesis stem

Item-Number: 172-916/12

Lot-Numbers: All after manufacturing date: 2013-10 

LINK® Reference-Number: R-2018-03

Dear Sir or Madame,

a risk has been identified for the use of the MP prosthesis stem. For safety reasons WALDEMAR LINK GmbH & Co. KG will recall these products. It is noticed that the smallest available stem size is not compatible with two corresponding instruments.

As you were supplied with the affected products, please read carefully the information on the following pages.

Please return the fax reply to us in any case until **12 December 2018**, even if you have none of the listed products in stock or if these products do not exhibit the defect in question.

In conclusion, we apologize for any inconveniences caused and thank you for your understanding in this matter.

Yours sincerely,
WALDEMAR LINK GmbH & Co. KG

Kai Lemke
Safety Officer